

S5 Table. Sensitivity analysis: false positive tests expected with varying RDT specificity.

Specificity	Peacekeepers barred from deployment due to false positive test, <i>n</i> of 454
99%	4 (1, 9)
95%	23 (14, 32)
90%	45 (33, 58)
80%	91 (74, 108)
70%	136 (117, 156)
60%	182 (161, 202)
50%	227 (206, 248)

Values are reported as median (95% CrI) based on quantiles of the corresponding Binomial distributions (S1 Text §2.2). Expected false positive outcomes did not differ across the range of incidence rates assessed (0.5/1000 PYAR to 10.0/1000 PYAR).