



**S1 Fig. Calibration of one peristaltic pump.** The calibration was made with measuring different volumes of water after having rotated the pump one minute at different velocities.